

AUG 5 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>C. R. Rasmussen</u>	ID# <u>3791</u> Date <u>8/5/10</u>
A Agency <u>APD</u> Phone # <u>786-2640</u>	
Instrument Location <u>Anchorage Jail</u>	
B Alco S/N <u>21</u> Target Value <u>.090</u> High Pressure <u>500</u>	
Alco Test Values <u>.078</u> <u>.078</u>	
1st Alco 2nd Alco	
Signature <u>C. R. Rasmussen</u> <u>R. Rasmussen</u> <u>1349</u> <u>CRB</u>	
8/10/10	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of AUGUST, 2010.

Daniel L. Nicholas
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130290

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number 130290 Model DataMaster
Location ANC1 Anchorage Police Dept.

CAB
8/5/10

Subject Information

Last Name Verification Of Calibration O.L. # ENRT TO CRIME LAB
First Name Radio Interference Out Of Service

Operator Information

Operator Name Ritale Case Z
Operator # 3791 Department ANC1

Test Results

Alco Target	0.080	Alco Serial	21
Nominal	0.079	Barometer	29.90
Dry Gas 1	0.078	Time	22:09:33
Subject Alcohol	0.000	Test Date	08/05/2010
Dry Gas 2	0.078	Time	22:10:46
		Test Time	22:08:03
		Report	Y

Diagnostic Test Record

Serial Number 130290 Model DataMaster
Location ANC1 Anchorage Police Dept.

Diagnostic Test Results

Test Date	08/05/2010	Computer	Okay
Test Time	22:14:44	Program	Okay
		Software Date	01/08/2009
		Sample Chamber (C)	48
		Breath Tube (C)	41
		Barometer (in)	29.88
		Flow Detector	Okay
		High Speed Pump	Okay
		Detector	Okay
		Filter	Okay
		Quartz	Okay
		Calibration	Okay

ALCO ADJUSTED FOR 29.88 in

ALCO TARGET	0.079	08/05/10
BLANK TEST	0.000	08/05/10
INTERNAL STANDARD	VERIFIED	08/05/10
ALCO IV 29.88 in	0.079	08/05/10
BLANK TEST	0.000	08/05/10
SUBJECT SAMPLE	0.000	08/05/10
BLANK TEST	0.000	08/05/10
ALCO IV 29.88 in	0.079	08/05/10
BLANK TEST	0.000	08/05/10

CAB
8/5/10
CR1349
8/5/10

HIGH SPEED	OKAY
DETECTOR	OKAY
FILTERS	OKAY
QUARTZ STANDARD	OKAY
CALIBRATION	OKAY

PRINTER TEST
OKAY
ALCO ADJUSTED FOR 29.88 in
08/05/10

CR1349
8/5/10